

CALIBRATION CERTIFICATE

Scale is calibrated@26°C/48%RH

It is certified that the below instrument was calibrated using procedures and calibration Equipment traceable to the National Instrument of Standards & Technology (NIST).

CUSTOMER GasQuip	MODEL GCS	G4 PART NUMBER S6000-000-01	CAPACITY 500lb
SERIAL NUMBER 2180849	PO NUMBER PO18405	TEST DATE 5/29/2025	SCALE ERROR -0.1lb

All readings are in lb

LINEARITY TEST					
Load	READINGS				ERROR
	RUN 1	RUN 2	RUN 3	AVERAGE	
0.0	0.0	0.0	0.0	0.0	0.0
50.0	49.9	49.9	49.9	49.9	-0.1
100.0	99.9	99.9	99.9	99.9	-0.1
150.0	149.9	149.9	149.9	149.9	-0.1
200.0	199.9	199.9	199.9	199.9	-0.1
250.0	249.9	249.9	249.9	249.9	-0.1

REPEATABILITY TESTS		
LOAD	READINGS	ERROR
240.0	239.9	-0.1
240.0	239.9	-0.1
240.0	239.9	-0.1
240.0	239.9	-0.1
240.0	239.9	-0.1
240.0	239.9	-0.1

INDICATOR	ECCENTRICITY TEST (40.0)		BATTERY
	40.0	39.9	
	39.9	39.9	